

## Declaration of Conformity

According to ISO/IEC 17050-1:2010

### We as a manufacturer:

HK Instruments Oy  
Keihästie 7  
FIN-40950 MUURAME, Finland  
[www.hkinstruments.fi](http://www.hkinstruments.fi)

### Declare that the products:

DPT250 - R8 (-AZ -D -Q -S)  
DPT2500 - R8 (-AZ -D -Q)  
DPT7000 - R8 (-AZ -D -Q)

### Conform to the following Directives:

EMC Directive 2014/30/EU  
RoHS Directive 2011/65/EU  
WEEE Directive: 2012/19/EU  
Product Safety: 2001/95/EC

### According to the following relevant standards:

IEC/EN61010-1: 2001 (Second edition) Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements

EN 61326-1: 2006 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

EN 61326-2-3: 2006: Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-3: Particular requirements - Test configuration, operational conditions and performance criteria for transducers with integrated or remote signal conditioning

### According to the following EMC test methods:

EN 61000-4-2: 2001  
EN 61000-4-3: 2002  
EN 61000-4-4: 2004  
EN 61000-4-5: 2001  
EN 61000-4-6: 2001

### Supplementary information:

These products comply with the requirements of the EMC Directive 2014/30/EU, and carries the CE-marking accordingly.

Muurame, June 2016

  
Tuomo Kalliomäki  
CEO